Capello Patient Positioner

Designed by William Capello, MD

Provides stable positioning of a patient during hip procedures

Board is available in a one-piece or two-piece design.
Optional two-piece board construction allows for easier use and storage.
All gel pads, pegs and peg height extensions can be used with existing peg boards.

PRODUCT NO'S:	
4090	[Set with 2-Piece Board and Standard Table Clamp]
4090-EU	[Set with 2-Piece Board and EU Table Clamp]
4095	[Set with 1-Piece Board and Standard Table Clamp]
4095-EU	[Set with 1-Piece Board and EU Table Clamp]]
Replacement Parts:	
4090-PB	[2-Piece Positioning Board]
4095-PB	[1-Piece Positioning Board]
4090-06	[6" (15,2 cm) Radiolucent Peg] *Four included in set; one with this product number*
4090-08	[9" (22,9 cm) Radiolucent Peg] *Four included in set; one with this product number*
4090-SC	[Stabilizing Clamp] *Two included in set; one with this product number*
4090-01	[Large Gel Pad]
9120	[Table Clamp] *Two included in set; one with this product number*
9120-EU	[Table Clamp for European Bed Rails] *Two included in set; one with this product number*
Optional Parts:	
4090-02	[Peg Gel Pad]
4090-EXT2	[Peg Extension – 2" (5,1 cm)]
4090-EXT	[Peg Extension – 4" (10,2 cm)]
4090-EXT6	[Peg Extension – 6" (15,2 cm)]
4090-EXT8	[Peg Extension – 8" (20,3 cm]

Each set includes:
Board, Gel Pad, (4) 6" Radiolucent Pegs, (4) 9" Radiolucent Pegs, (2) Stabilizing Clamps, (2) Table Clamps.

USA MADE

Optional Peg Gel Pad

Optional 2", 4", 6", & 8" Peg Extensions

Optional two-piece board design for easy handling (Also available in a one-piece design)

Board Dimensions: 47" x 18.75" (120 cm x 47,6 cm)

– Over –

www.innomed.net
FREE TRIAL ON MOST INSTRUMENTS
INNOMED®
ISO 13485:2016
© 2026 Innomed, Inc.

103 Estus Drive, Savannah, GA 31404
info@innomed.net
912.236.0000 Phone
912.236.7766 Fax
Innomed-Europe Tel. +41 41 740 67 74
info@innomed-europe.com
1.800.548.2362

Capello Patient Positioner

Set-up guide for patient positioning in total hip surgery



Step 1:
The positioner is placed on the surgical table and secured with the two stabilizing clamps provided – one at the head of the table, and the other at the foot of the positioner on the opposite side of the table. The clamps are secured and locked to the table rails using the two table clamps.



Step 2:
The gel pad is placed on the positioner. A sheet (not shown) can then placed over the padding to help prevent shifting as the patient is transferred to the table.





Step 3:
Radiolucent pegs should be positioned as follows:

1) A short peg placed in the sacral area. A short peg is used here instead of a long peg in order to maximize the operative area.
2) A short peg placed in front of the pubis. Care should be taken to avoid neurovascular structure compression in both femoral triangles.
3) A long peg placed in the posterior thoracic region below the scapula.
4) A long peg placed directly below the pectoral muscle, keeping in mind the need for adequate ventilation of this area.

Note: More radiolucent pegs, as well as the use of the peg extenders, may be required for larger patients.



CAUTION: The board should not be used as a transport device.

– Over –

Innomed, a manufacturer of surgical instruments, does not practice medicine and does not recommend this or any other surgical technique for use on a specific patient. The surgeon who performs any procedure is responsible for determining and utilizing the appropriate techniques for such procedure for each individual patient. Innomed is not responsible for the selection of the appropriate surgical technique to be used for an individual patient.